

Temperature Log

Month & Year _____

Provider PIN _____

Storage Unit Location/ID	Scale ☐ Fahrenheit ☐ Celsius	Storage Unit Type (select one) ☐ Refrigerator (36° to 46°F / 2° to 8°C) ☐ Freezer (-58° to 5°F / -50° to -15°C) ☐ ULT (-130° to -76°F / -90° to -60°C)
Funding Source(s) ☐ VFC ☐ VFA ☐ LHD 317 ☐ SGF ☐ Private ☐ _____		

Day	Time	Initials	Alarm	Current	Min	Max	Excursion #
Example	8:00 am	NN		40.5	38.1	43.7	
	4:00 pm	NN	✓	37.4	33.0	39.2	12345
1	am						
	pm						
2	am						
	pm						
3	am						
	pm						
4	am						
	pm						
5	am						
	pm						
6	am						
	pm						
7	am						
	pm						
8	am						
	pm						
9	am						
	pm						
10	am						
	pm						
11	am						
	pm						
12	am						
	pm						
13	am						
	pm						
14	am						
	pm						
15	am						
	pm						

Check Temperatures Twice a Day:

1. Fill out clinic/unit details in header.
2. Record the time and your initials.
3. Record a check if alarm went off.
4. Record Current, Min, and Max.
5. Clear Min/Max on your device.
6. Ensure data logger is recording.

If Alarm Went Off:

1. Clear Min/Max and alarm symbol.
2. Post "Do Not Use Vaccines" sign.
3. Alert your supervisor.
4. Report excursion at myCAvax for all funding sources.
5. Record myCAvax Batch Excursion # in last column or add notes. Ensure data logger is recording.

On-Site Supervisor's Review

When complete, check all that apply:

- ☐ Temperatures were recorded twice daily.
- ☐ I reviewed data files to find any missed excursions.
Download date: _____
- ☐ Any excursions were reported.

On-Site Supervisor's Name: _____

Signature: _____

Date: _____

Staff Names and Initials:

Keep all temperature logs and data files for three years.

Falsifying log is grounds for vaccine replacement and program termination.

VFC: (877) 243-8832

VFA, LHD 317, SGF: (833) 502-1245

Temperature Log

Month & Year _____

Provider PIN _____

Storage Unit Location/ID	Scale ☐ Fahrenheit ☐ Celsius	Storage Unit Type (select one) ☐ Refrigerator (36° to 46°F / 2° to 8°C) ☐ Freezer (-58° to 5°F / -50° to -15°C) ☐ ULT (-130° to -76°F / -90° to -60°C)
Funding Source(s) ☐ VFC ☐ VFA ☐ LHD 317 ☐ SGF ☐ Private ☐ _____		

Day	Time	Initials	Alarm	Current	Min	Max	Excursion #
16	am pm						
17	am pm						
18	am pm						
19	am pm						
20	am pm						
21	am pm						
22	am pm						
23	am pm						
24	am pm						
25	am pm						
26	am pm						
27	am pm						
28	am pm						
29	am pm						
30	am pm						
31	am pm						

Check Temperatures Twice a Day:

1. Fill out clinic/unit details in header.
2. Record the time and your initials.
3. Record a check if alarm went off.
4. Record Current, Min, and Max.
5. Clear Min/Max on your device.
6. Ensure data logger is recording.

If Alarm Went Off:

1. Clear Min/Max and alarm symbol.
2. Post "Do Not Use Vaccines" sign.
3. Alert your supervisor.
4. Report excursion at myCAvax for all funding sources.
5. Record myCAvax Batch Excursion # in last column or add notes. Ensure data logger is recording.

On-Site Supervisor's Review

When complete, check all that apply:

- ☐ Temperatures were recorded twice daily.
- ☐ I reviewed data files to find any missed excursions.
Download date: _____
- ☐ Any excursions were reported.

On-Site Supervisor's Name: _____

Signature: _____

Date: _____

Staff Names and Initials:

Keep all temperature logs and data files for three years.
Falsifying log is grounds for vaccine replacement and program termination.

VFC: (877) 243-8832

VFA, LHD 317, SGF: (833) 502-1245